

Application of screen-printed carbon electrode as an electrochemical transducer in biosensors

Abstract

Screen printed electrodes have been extensively employed as an economical transducer substrate for electrochemical biosensing applications due to their small size, easiness of mass production and the possibility of use with portable devices which facilitates *in situ* applications. Carbon inks can be widely modified by the addition of materials and/or molecules and this versatility confers the capacity to be used for the purposes of food, agricultural, environmental and biomedical analyses.

Keywords: screen printed electrode, carbon, biosensing

Volume 5 Issue 1 - 2019

Ricardo Adriano Dorledo de Faria,^{1,2}Younès Messaddeq,^{2,3} Luiz Guilherme Dias Heneine,⁴ Tulio Matencio⁵

¹Department of Chemical Engineering, Universidade Federal de Minas Gerais, Brazil

²Center for Optics, Photonics and Lasers (COPL), Université Laval, Canada

³Institute of Chemistry, Brazil

⁴Department of Applied Immunology, Fundação Ezequiel Dias, Brazil

⁵Department of Chemistry, Universidade Federal de Minas Gerais, Brazil

Correspondence: Ricardo Adriano Dorledo de Faria, Center for Optics, Photonics and Lasers (COPL), Pavillon d'optique-photonique, Université Laval, 2375 rue de la Terrasse, Quebec, Canada, Tel (418) 656 2131 (Ext 7184), Email ricardo.adriano08@hotmail.com

Received: December 18, 2018 | **Published:** January 02, 2019

Introduction

Great interest has been directed to the screen printing technology for biosensing applications. Screen-printed electrodes (SPEs) possess as major advantage over the traditional electrodes the possibility of use in miniaturized systems, whose applications demand portable devices. Besides the possibility to perform *in situ* analyses, the easiness of mass production makes these electrodes very interesting for the market.

Fabrication

Usually, the SPEs consist of a three-electrode configuration (working, counter- and reference electrodes). Carbon ink is painted onto one extremity of the conductive tracks to form the working and counter electrodes, and the reference electrode is commonly based on silver electroactive tracts. An insulating film is glued to the exposed tracts and between the electrodes to avoid short-circuit. The screen-printed carbon electrodes (SPCEs) contain a carbon conductive ink that is chemically inert, organic solvents, additives and some binding components.^{1,2}

Pre-treatment

Since the SPCE possess insulator additives to improve the adhesion of carbon ink on the support, the pre-treatment of these electrodes is considered a key point to overcome its limited electron transfer kinetics at the interface with the electrolyte. In order to activate the edge planes of the SPCE, many techniques have been considered to enhance the carbon electro activity (i.g. thermal, chemical and mechanical treatments), although the electrochemical processes are the most commonly used. Sundaresan et al.,³ performed 10 cycles of cyclic voltammetry in 0.05 M phosphate buffer in a potential range from -0.5 to 2.0 V vs Ag/AgCl. Pan et al.,⁴ applied a fixed -1.2 V vs Ag/AgCl potential during 20 s to an electrode containing a drop of 0.1M NaOH.

Applications in electrochemical biosensing

The final step to produce an SPCE fit for purpose is the modification (functionalization) of the working electrode surface. The target analyte drives the choice of molecules or biomolecules with specific reactivity towards it to be attached. Table 1 shows some recent applications of SPCE regarding the biosensing of molecules of medical, environmental and food interests.

Table 1 Recent applications of SPCE as electrochemical transducer in biosensors

Application	Target analyte	Molecule of recognition	Technique of detection	Reference
Medical diagnosis	Cardiac troponin I Japanese encephalitis virus (JEV)	5'-amine modified Tro4 aptamer Anti-JEV antibodies	Chronoamperometry Electrochemical Impedance Spectroscopy	5,6
Environment monitoring	Cathecol Bisphenol A	Laccase Dendritic platinum NPs	Chronopotentiometry Cyclic voltammetry	7,8
Food analysis	Fructose Ethyl carbamate	Graphite NPs O-aminophenol	Chronoamperometry Cyclic voltammetry	9,10

Conclusion

The possibility of modification of SPCE to detect various molecules associated to its miniaturized dimensions and low cost of production has been extensively exploited, as reported in the literature, for the development of versatile electrochemical biosensors. The electro activity of carbon has been improved by different treatments making it suitable for application as a transducer in electroanalysis mainly towards the detection of biological molecules.

Acknowledgements

The authors thank to the Canada Excellence Research Chair in Photonics Innovations (CERCPI) and the Natural Sciences and Engineering Research Council of Canada (CRSNG).

Conflicts of interest

There are no conflicts of interest.

References

1. González-Sánchez MI, Gómez-Monedero B, Agrisuelas J, et al. Highly activated screen-printed carbon electrodes by electrochemical treatment with hydrogen peroxide. *Electrochem Commun.* 2018;91:36–40.
2. Tudorache M, Bala C. Biosensors based on screen-printing technology, and their applications in environmental and food analysis. *Anal Bioanal Chem.* 2007;388(3):565–578.
3. Sundaresan P, Chen TW, Chen SM, et al. Electrochemical Activation of Screen Printed Carbon Electrode for the Determination of Antibiotic Drug Metronidazole. *Int J Electrochem Sci.* 2018;13:1441–1451.
4. Pan D, Rong S, Zhang G, et al. Amperometric determination of dopamine using activated screen-printed carbon electrodes. *Electrochem.* 2015;83(9):725–729.
5. Jo H, Her J, Lee H, et al. Highly sensitive amperometric detection of cardiac troponin I using sandwich aptamers and screen-printed carbon electrodes. *Talanta.* 2017;165:442–448.
6. Chin SK, Lim LS, Pang SC, et al. Carbon nanoparticle modified screen printed carbon electrode as a disposable electrochemical immunosensor strip for the detection of Japanese encephalitis virus. *Microchim Acta.* 2016;184(2):491–497.
7. Palanisamy S, Ramaraj SK, Chen SM, et al. A novel Laccase Biosensor based on Laccase immobilized Graphene-Cellulose Microfiber Composite modified Screen-Printed Carbon Electrode for Sensitive Determination of Catechol. *Nature.* 2017;7:1–12.
8. Shim K, Kim J, Shahabuddin M, et al. Efficient wide range electrochemical bisphenol-A sensor by self-supported dendritic platinum nanoparticles on screen-printed carbon electrode. *Sens Actuat B.* 2018;255:2800–2808.
9. Nicholas P, Pittson R, Hart JP. Development of a simple, low cost chronoamperometric assay for fructose based on a commercial graphite-nanoparticle modified screen-printed carbon electrode. *Food Chem.* 2018;241:122–126.
10. Zhao X, Zuo J, Qiu S, et al. Reduced Graphene Oxide-Modified Screen-Printed Carbon (rGO-SPCE)-Based Disposable Electrochemical Sensor for Sensitive and Selective Determination of Ethyl Carbamate. *Food Anal Methods.* 2017;10(10):3329–3337.